

ROY COOPER  
Governor  
DIONNE DELLA-GATTI  
Secretary  
MICHAEL A. ABRACZINSKAS  
Director



## DAQ-08-002.1 Standard Operating Procedure (SOP) Model T500U Nitrogen Dioxide (NO<sub>2</sub>) Monitoring System Electronics and Calibration Branch (ECB) Responsibilities




## 1.0 Approval Sign Off-Sheet

---

I certify that I have read and approve of the contents of the Model T500U Nitrogen Dioxide (NO<sub>2</sub>) Monitoring System – 200 PPB Electronics and Calibration Branch (ECB) Responsibilities Standard Operating Procedure with an effective date of April 23, 2021.

### Director, Air Quality Division

Michael Abraczinskas

Signature:  Date: 4/19/2021  
DS013FED00218423...

### Ambient Monitoring Section Chief

Patrick Butler

Signature:  Date: 4/19/2021  
4290A7E2D2A0420...

### Laboratory Analysis Branch Supervisor

Jim Bowyer

Signature:  Date: 4/19/2021  
G96894826C9B41D...

### Projects and Procedures Branch Supervisor

Joette Steger

Signature:  Date: 4/16/2021  
B278F8186CEC4DC...

### Primary SOP Author

Kay Roberts

Signature:  Date: 4/16/2021  
BE4009EE242941D...

This document, and any revision hereto, is intended solely as a reference guide to assist Electronics and Calibration Branch technicians in the setup, calibration, maintenance, operation, and the collection of data related to the North Carolina Division of Air Quality's Ambient Monitoring Program. This document is intended as a supplement to, and not a substitute for, the education, training, and experience in addition to other North Carolina DAQ Reference Manuals, QAPPs and EPA Guidance. It is not a substitute for the extensive training and experience required for the efficient operation of ambient air monitoring equipment needed for scientifically valid data.

## Table of Contents

---

1.0	Approval Sign Off-Sheet.....	2
2.0	SCOPE AND PURPOSE .....	5
3.0	EQUIPMENT SELECTION AND PROCUREMENT .....	5
3.1	Equipment and Material List.....	5
4.0	TELEDYNE MODEL T500U DESCRIPTION AND SPECIFICATIONS (from Manual).....	6
5.0	DESCRIPTION OF EQUIPMENT AND CHECKS .....	7
5.1	Teledyne T500U CAPS Monitor Testing .....	7
5.2	Teledyne T700U Dynamic Gas Calibrator .....	8
5.3	Zero Air Generator and Certification .....	8
5.3.1	Zero Air Generator Checks.....	9
5.3	EPA Protocol Gas Cylinder .....	9
5.4	Dedicated Site Computer and Modem with a Data Acquisition System .....	10
5.5	Temperature Controlled Monitoring Shelter.....	10
6.0	INITIAL LABORATORY SET-UP .....	10
6.1	Inspection .....	11
6.2	Initial Laboratory Set-up .....	11
6.2.1.	Calibrator Set-up.....	11
6.2.1.	T500U Set-up .....	11
6.3	Diagnostic Checks and Setting .....	12
6.4	Monitor Filters .....	12
6.4	Leak and Flow Tests .....	12
6.4.1.	Detailed Pressure Leak Check .....	13
6.4.2.	Sample Flow Check .....	13
7.0	ON-SITE INSTALLATION.....	13
7.1	Monitoring Equipment Site Installation.....	13
	Installation at the monitoring site includes: .....	13
7.2	Siting Criteria.....	14
7.3	Site Computer, DAS and Modem .....	14
7.4	Leak and Flow Checks .....	14
7.4.1.	Detailed Pressure Leak Check .....	14
7.4.2	Sample Flow Check .....	14
7.5	Diagnostic Checks upon Installation .....	14
8.0	ROUTINE MAINTENANCE.....	15

8.1 Teledyne T500U NO<sub>2</sub> Analyzer .....15

8.2 Teledyne Model T700U Calibrator.....15

8.3 Model 701 Zero Air Generator Checks .....15

9.0 ACCURACY AUDITS.....16

9.1 Audit Procedures .....16

10.0 REVISION HISTORY .....17

11.0 REFERENCES.....17

12.0 APPENDICES .....17

## 2.0 SCOPE AND PURPOSE

---

The Electronics and Calibration Branch (ECB) of the Ambient Monitoring Section (AMS) of the Division of Air Quality (DAQ) is responsible for the selection, evaluation, and procurement of the NO<sub>2</sub> monitoring equipment and related accessories. Further ECB is responsible for the receipt, assembly, testing (at its facility) and installation of NO<sub>2</sub> monitors in the field, evaluation of the on-going performances of NO<sub>2</sub> monitors and related support equipment and scheduled and unscheduled maintenance of the system. As part of its responsibilities, ECB is also expected to maintain a sufficient inventory of monitors, support equipment and replacement parts to minimize loss of NO<sub>2</sub> ambient monitoring data.

Additionally, ECB staff is also responsible for procuring and maintaining dedicated traceable nitric oxide (NO) standards for the certification of all calibrators and for the independent accuracy auditing of the ambient air quality NO<sub>2</sub> monitors. These standards provide a direct link to established national standards and thus become the basis for the collection of the highest quality ambient monitoring NO<sub>2</sub> data in accordance with current procedures and existing Federal Regulations and Guidelines. The continual accuracy audits performed by the ECB staff provide an ongoing evaluation of the NO<sub>2</sub> monitor's performance and site operator's adherence to DAQ approved operating procedures.

The ECB also maintains permanent records of all NO standards used in the calibration and auditing of monitors and sampling equipment used in support of DAQ monitoring activities. There are permanent records at ECB for each NO<sub>2</sub> monitor and samplers used to analyze ambient air quality in the State of North Carolina. Each major component of the NO<sub>2</sub> monitoring system, such as analyzer, calibrator, zero-air supply system, etc., is assigned a dedicated logbook. These logbook records include information relate to the performance evaluations and complete records detailing the instruments and equipment placed at each monitoring site. Both permanent records are updated continuously.

The ECB is also responsible for evaluating, developing, and recommending changes in the equipment and operating parameters to improve the quality of data collected and procedures used in the collection of data.

## 3.0 EQUIPMENT SELECTION AND PROCUREMENT

---

The North Carolina Ambient Air NO<sub>2</sub> Monitoring System must meet or exceed the Reference and Equivalent Method requirements in 40 CFR Part 53.1 and 40 CFR Part 58, Appendix C. The North Carolina ambient NO<sub>2</sub> monitoring system consists of the items listed in Section 3.1, below. Minor components are not specified but included by reference.

### 3.1 Equipment and Material List

---

- Teledyne Model T500U NO<sub>2</sub> Monitor
- Teledyne – API Model T700U Dynamic Gas Calibrator
- Teledyne – API Model 701 Zero Air Generator
- EPA Protocol NO gas cylinder
- Teflon™ Sampling Line
- Data Management System (e.g., Envidas/Envista)
- Dedicated Windows compatible Site Computer
- Ethernet / Modem connectivity
- Temperature Controlled Monitoring Shelter

The ECB is responsible for ensuring that all components are compatible with the measurement of ambient levels of atmospheric NO<sub>2</sub>. The ECB is responsible for the performance of complete system evaluation prior to the field installation and that the system is fully functional at the completion of the installation. On an ongoing basis as needed the ECB provides equipment and instrumentation maintenance and operational support to maximize the collection of the highest quality ambient air pollution data possible in accordance with accepted and approved procedures.

#### 4.0 TELEDYNE MODEL T500U DESCRIPTION AND SPECIFICATIONS (from Manual)


The Teledyne Model T500U CAPS NO<sub>2</sub> Analyzer uses Cavity-Attenuated Phase-Shift (CAPS) spectroscopy to render true measurements of NO<sub>2</sub>. The T500U operates as an optical absorption spectrometer, wherein the absorbance (lost light) is directly proportional to the path length and the concentration of the absorbing gas (Beer-Lambert law), providing direct measurement of NO<sub>2</sub>.

The T500U uses few components: an optical cell, a pair of highly reflective spherical mirrors centered at 450 nanometers (strong NO<sub>2</sub> absorbance band), a light emitting diode (LED), and a vacuum photodiode detector behind the other mirror at the opposite end of the cell. The LED emits ultraviolet (UV) light into the cell; the light reflects back and forth between the two mirrors, building intensity and running a very long path length. The long path length extends the “time” or “life” of the photon, thus providing ample time to measure absorbance when NO<sub>2</sub> is present. Through the use of precisely timed data acquisition coupled with a proprietary algorithm the measure absorption is translated into a phase shift, from which the NO<sub>2</sub> concentration is calculated. The phase shift decreases as the NO<sub>2</sub> signal increases.

The CAPS method is faster than the traditional chemiluminescence method since the sample does not require cycling through a catalytic converter to calculate a difference measurement. Its speed also makes measurement more precise due to the ability to capture samples closer to “real time” before ventilation vortices (e.g., urban canyons and other traffic-related forces) can scatter the concentration.



Teledyne Model T500U NO<sub>2</sub> analyzer front panel display



T500U Pneumatics with Internal Zero Span Values Option

## 5.0 DESCRIPTION OF EQUIPMENT AND CHECKS

### 5.1 Teledyne T500U CAPS Monitor Testing

The CAPS monitor should be tested thoroughly before deployment at the monitoring site. This testing should include but is not limited to:

- Pre-calibration electronic adjustment
- Data Acquisition output adjustments
- Setting initial calibration factors and adjustments
- NO<sub>2</sub> operational test
  - Zero / Span / Precision Sequence
  - Calibration phases
  - Verification Sequence

Only the main components of the CAPS NO<sub>2</sub> monitoring system are discussed briefly here for their operation details. For further details of other NO<sub>2</sub> monitoring system related components refer to the "Model T500U CAPS NO<sub>2</sub> Analyzer, Operation Manual, 07834D DCN6998, 01 Jun 2015.

## 5.2 Teledyne T700U Dynamic Gas Calibrator

The T500U analyzer is calibrated using a T700U calibrator, which must have flows certified by ECB and traceable to a primary standard according to the requirements in the QA/SOP Gas Calibrators (2.3.4 R12.2). The three mass flow controllers and ozone (O<sub>3</sub>) photometer in the T700U “audit calibrator” is certified for twelve (12) months and the “field calibrator” certifications are good for twelve (12) months.

The calibrator is an accurate mass flow-controlled gas dilution system that meets the 40 CFR Part 50 requirements of  $\pm 2\%$  accuracy. “NO<sub>2</sub>” gas (usually in an inert gas such as nitrogen) from a National Institute of Standards and Technology (NIST) traceable protocol certified cylinder (of  $\pm 2\%$  accuracy or less) is blended with “zero-air” to provide the desired concentrations. From the known calibration of the three mass flow controllers, the exact concentration can be calculated. The typical dilution ratio of 100:1 to 1000:1 is generally used to generate appropriate concentrations.

The Model T700U dynamic gas calibrator supplies the required concentration of “NO” to perform Zero, Precision, Span Checks, Calibrations and Multi-point Verification by gas phase titration (GPT).

The principle of GPT is based on the rapid gas phase reaction between NO and O<sub>3</sub>, which produces quantities of NO<sub>2</sub> as shown by the following equation:



It has been empirically determined that under controlled circumstances the NO-O<sub>3</sub> reaction is very efficient (<1% residual O<sub>3</sub>), therefore the concentration of NO<sub>2</sub> resulting from the mixing of NO and O<sub>3</sub> can be precisely predicted and controlled as long as the following conditions are met:

- The amount of O<sub>3</sub> used in the mixture is known;
- The amount of NO used in the mixture is *at least* 10% greater than the amount of O<sub>3</sub> in the mixture;
- The volume of the mixing chamber is known; and
- The NO and O<sub>3</sub> flow rates (from which the time of the reactants in the mixing chamber) are low enough to give a residence time of the reactants in the mixing chamber of >2.75 parts per million (ppm) min.

Given the above conditions the amount of NO<sub>2</sub> being output by the T700U calibrator will be equal to (at a 1:1 ratio) to the amount of O<sub>3</sub> added.

The O<sub>3</sub> flow rate of the T700U O<sub>3</sub> generator is a set fixed value (typically about 0.105 liters per minute) and the GPT chamber’s volume is known. Thus, once the *total gas flow* requirements, the source concentration of NO, and the target concentration for the O<sub>3</sub> generator are entered into the calibrator’s software, the T700U adjusts the NO flow rate and diluent (zero air) flow rate to create the appropriate NO<sub>2</sub> concentrations at the output.

## 5.3 Zero Air Generator and Certification

The Teledyne Series 701 Zero Air Generator is a pure air generator system that is capable of continuous delivery of up to 20 standard liters per minute (SLPM), 30 pounds per square inch (PSI) of dry, contaminant-free air. The air is suitable for use as: a zero-reference calibration gas, ultra-pure combustion air for flame ionization detector, and service air for pneumatically operated valves. The system can deliver air free from water vapor, particulates, sulfur dioxide (SO<sub>2</sub>), Hydrogen Sulfide (H<sub>2</sub>S), NO, NO<sub>2</sub>, Ozone (O<sub>3</sub>), and Carbon Monoxide (CO). The delivery pressure should be set to  $30 \pm 2$  pounds per square inch gauge (psig).

The Teledyne API Model 701 Zero Air Generator system is a source of clean, dry purged air. The 701 produces scrubbed air via an oil and diaphragm free pump which removes SO<sub>2</sub>, NO, NO<sub>2</sub>, O<sub>3</sub> and H<sub>2</sub>S. Inlet air is pulled through a water trap and pre-cooler to remove moisture. The scrubbed air then passes through a regenerative



dryer for final drying operations. Lastly, outlet air then passes through a filter to assure clean, dry, zero air with a dew point of less than -20 degrees Celsius (°C) independent of the inlet dew point.

### 5.3.1 Zero Air Generator Checks

Model 701 Zero Air Generator Checks:

- The pollution scrubber/converter media should be replaced annually by the ECB.
- Verify that the delivery pressure is set to  $30 \pm 2$  psig. (If the delivery pressure is outside of the  $\pm 2$  psig range, adjust the pressure using pressure adjust control knob.)
- Check the drain from the air generator for condensation.

### 5.3 EPA Protocol Gas Cylinder

The ECB shall procure certified EPA Protocol procedure G1 gas cylinders for the Ambient Monitoring Section. Primary “NO” standards are used when calibrating and during the nightly calibration checks, manual performance checks and annual performance audit of the CAPS NO<sub>2</sub> monitors. The primary “NO” standard used must be a certified, commercially prepared compressed gas standard with a certified accuracy of no worse than  $\pm 2$  percent.

Standards in the concentration range of  $\sim 10$  ppm are suitable choices for dilution to prepare low concentration mixtures.

- Extreme care must be taken to ensure compatibility for all components. Flow rates and concentration outputs must meet the requirements of the monitor.
- All EPA Protocol G1 calibration gas standards must be reference to NIST nitrogen in Air Standard Reference Material (CRM). A written statement of certification should be obtained which provides the following:
  1. A brief description of the certification procedure,
  2. Cylinder numbers,
  3. Cylinder gas concentrations,
  4. Replicate analysis data,
  5. Balance gas used
  6. NIST, SRM numbers used as standards, and
  7. Last analysis date.

A copy of this certification should be available to users and will be kept on file in the ECB files

- The calibration gas standards will have their own certifications and will be considered as expired after 3 years.
- NO cylinder gas should not be used below a cylinder pressure of 500 psig as shown by the cylinder gas regulator.
- Each “NO” span gas cylinder shall contain the following minimum traceability information on a label or tag affixed to the cylinder or valve:
  1. The concentration of cylinder gas
  2. The last analysis date,
  3. The expiration date,
  4. The initials of the person performing the analysis,
  5. Cylinder number, and
  6. Balance gas.

#### 5.4 Dedicated Site Computer and Modem with a Data Acquisition System

The ECB is responsible for providing dedicated instrumentation and operating system compatible computer, Data Acquisition System (DAS), and a modem communication system. The operating system, data collection and storage capability, and remote communication system must be designed and acquired to maximize ease of data collection and transferring of data remotely. The entire data collection and communication system must be compatible with data retrieval programs operated by the database manager of the Ambient Monitoring Section. The ECB is encouraged to evaluate procedures, equipment, and systems to improve the data collection, data storage and data transferring capabilities of the NC Ambient Monitoring system.

#### 5.5 Temperature Controlled Monitoring Shelter

Each CAPS monitoring site shelter must be maintained in good repair. Each shelter must be temperature controlled within the EPA requirements of 5.0 – 40.0 °C with a daily fluctuation less than 2.1 °C standard deviation over a 24-hour period, if the CAPS is in its own shelter, or within 20.0 to 30.0 °C, if needed for other monitors. The temperature is monitored and recorded every minute through Envidas via the Comet temperature probe/sensor. The back-up temperature recording system is the HoBo temperature sensor and must be manually downloaded by the site operator at least once per month. Also, a certified mineral thermometer is kept at each site and serves as the primary standard to compare the Comet and HoBo values to assure accuracy. At installation, the ECB verified the temperature with a standard calibrated and certified thermometer (Omega RTD) and assures that the Comet and Hobo sensor display agree with the standard. The hourly site temperature can be viewed by the site operator, Regional Monitoring Coordinator, and the Central Office Staff, and quality assurance is confirmed daily to assure no data will be lost because of the site temperature drifting. During the annual performance evaluation audits, the ECB staff again checks the temperature against a standard verified and certified thermometer. If drift or deviation occurs outside of the 5.0 – 40.0 °C range, then the data are disqualified, and a void code (AE) is submitted in the daily data by the operator / reviewer to invalidate any potential erroneous equipment operation because of a temperature drift. In addition, the monitoring shelter provides a secure hazard proof facility for both instrumentation and monitoring personnel.

### 6.0 INITIAL LABORATORY SET-UP

---

The ECB shall conduct operational tests after receipt and unpacking of each instrument. Following the Teledyne T500U Instruction Manual (Sections 3 - 10) and Section 3.0 (Equipment Checks) and Section 5.0 (Detailed Procedures) of NC DAQ-08-001.2 Operator's SOP. The instrument must sample calibration gas at atmospheric pressure. After initial set-up and instrument checks, the instrument is either approved or returned to the manufacturer if any damage or problems that cannot be fixed are identified.

Upon approval of the tested unit, the unit shall be added to the fixed asset system. For each monitor, apply an inventory decal and complete an inventory load sheet showing the planned monitor location. Submit the inventory load sheet to the branch supervisor.

Prior to deployment in the field, all instruments will undergo basic operational tests in the ECB lab, with results recorded in the instrument's logbook which is filed in the ECB lab. (See Section 5.0 of the T500U Instruction Manual) The instrument should be set-up with accompanying calibrator, zero air generator, cylinder and data logging system.

## 6.1 Inspection

Visually inspect the exterior of all items for damage. Remove the cover and inspect the electronics assembly and circuit boards for loose wires, cables, broken components, or other damage. Reconnect any loose components and if necessary, contact the manufacturer.

## 6.2 Initial Laboratory Set-up

Prior to deployment in the field, all instruments will undergo basic operational tests in the ECB lab with results recorded in the instruments logbook which is filed in the ECB lab.


### 6.2.1. Calibrator Set-up

- Attach a vented Teflon™ tube (FEP Teflon type only) from the fitting labeled “CalGas Out” on the rear panel of the calibrator to the “Span and Zero Air” input of the monitor. Connect a source of zero-air to the inlet port labeled “Diluent In”. Connect the standard NO gas cylinder to the port labeled CYL 1.
- Assure the correct cylinder concentration is entered into the T700U. Bleed the cylinder.
- Plug in power cord and turn on power to instrument.

### 6.2.1. T500U Set-up

- Plug in power cord and turn on power to instrument.
- Connect the rear panel bulkhead labeled “Exhaust” to a suitable vent outside of the ECB lab or the monitoring room, *do not* vent to room air. (See diagram below.)
- Establish a connection and communication between the monitor, dedicated personal computer and the current DAS in use.
- Set the range of the monitor to 200 parts per billion (ppb).
- Using DAS software, create the necessary sequences and phases needed for operators to complete NO<sub>2</sub> operational test such as:
  - Zero / Span / Precision Sequences
  - Calibration phases
  - Verification Sequence

Each sequence and phase should be thoroughly tested prior to deployment of the NO<sub>2</sub> monitoring system at a site.



Teledyne Model T500U NO<sub>2</sub> analyzer rear panel display

### 6.3 Diagnostic Checks and Setting


The following are diagnostic menu set points for the T500U NO<sub>2</sub> analyzer:

Parameter	Expected Value
Instrument Range/Concentration Units	200 PPB
Sample Pressure (In Hg-A)	26 – 29 In Hg
AREF_L (Mm-1)	< 650 Mm-1 <sup>[1]</sup> while running Zero Air
Box Temperature (°C)	Ambient ± 5.0 °C
Oven Temperature (°C)	45 ± 1 °C
Phase (degree of angle)	> 25 degrees while running Zero Air
NO <sub>2</sub> Measured Loss	< 650 Mm-1 <sup>[2]</sup> while running Zero Air
Range Mode	Single
Slope #1	-0.8 to 1.2, Optimal is 1.000
NO2 Offset #1	± 10.0 ppb, Optimal is 0.0 ppb

<sup>[1]</sup> Mm-1: Unit of measurement to express light absorption, expressed in MEGA (10<sup>6</sup>) per meter

<sup>[2]</sup> Mm-1: Unit of measurement to express light absorption, expressed in ppb for NO<sub>2</sub>

The diagnostic menu is best accessed via the NumaView Dashboard:



### 6.4 Monitor Filters

The T500U operates with an internal in-line sample filter along with an Auto Reference (AREF) filter assembly (backup particle filter and charcoal filter, which is an NO<sub>2</sub> gas scrubber). The manufacturer recommends these filters be changed annually or when the number of system hours is between 8750 and 9000 hours, or more frequently if needed. If the AREF is observed to be increasing significantly or approaching values near 1100 Mm-1, ECB should be notified as this is an indicator that the monitor requires maintenance. Annual filter changes should be scheduled with and completed by ECB personnel.

### 6.4 Leak and Flow Tests

Leak and sample flow checks should be performed periodically as part of the standard maintenance of the T500U.

#### 6.4.1. Detailed Pressure Leak Check

Obtain a leak checker that has a small pump, shut-off valve, and pressure gauge to create an over-pressure condition, less than or equal to 10 psig. Alternatively, a tank of pressurized gas, with the two-stage regulator adjusted to  $\leq 10$  psig, a shutoff valve and a pressure gauge may be used. See section 11.3.5.1 of the manual for step-by-step instructions for completing the pressure leak check.

#### 6.4.2. Sample Flow Check

Sample flow checks are useful for monitoring the actual flow of the instrument. A decreasing sample flow may point to slowly clogging pneumatic paths, most likely the sample flow restrictor or the sample filter. See section 11.3.5.2 of the manual for step-by-step instructions for performing a sample flow check.

## 7.0 ON-SITE INSTALLATION

The ECB will install the monitor and its support equipment. Acquiring access to a site, and approval of the site is the responsibility of the DAQ regional office and the Projects and Procedures Branch. Internet / wireless service and electrical power should be secured along with any needed permits, new wiring, etc., prior to installation of the monitor and support equipment. The site location must meet the applicable site requirements and be approved by the Projects and Procedures Branch Supervisor and EPA.

Electrical circuits should be dedicated and properly sized prior to the installation of any monitoring equipment and auxiliary components. The site should be inspected for integrity and safety. The ECB staff is responsible for the installation of all State of North Carolina operated Teledyne T500U monitoring systems.

The monitor should be replaced on an as needed basis: the calibrators require recertification every twelve (12) months and should be replaced prior to their certification expiration date or as needed. The EPA Protocol Gas Cylinder should be replaced when the tank pressure is  $< 500$  psig or prior to the expiration date. All procedures should be documented on the 109 Form and all applicable site and instrument logbooks.

### 7.1 Monitoring Equipment Site Installation

Installation at the monitoring site includes:

- Teledyne Model T500U NO<sub>2</sub> monitor
- Teledyne Model T700U gas calibrator
- Teledyne Model 701 Zero Air Generator
- EPA Protocol NO gas cylinder
- Teflon™ tubing Sample Line with a Teflon™ inlet screen
- Data Management System (e.g., Envidas/Envista)
- Dedicated Windows compatible Site Computer
- Ethernet / Modem connectivity

The T500U monitor and auxiliary equipment must be installed in a building where room temperature extremes do not fall below 5.0 °C or exceed 40.0 °C. Check to ensure that the heater and air conditioner are in working order and do indeed maintain the desired temperature range, irrespective of the time of the day and season. Remove the air conditioner filter and clean if necessary. Check for any problems related to the building such as leaks, infestations, etc. Allow a warm-up period (~ 1 hour) for the monitor and auxiliary equipment after powering up before running any diagnostic test.

## 7.2 Siting Criteria

To ensure the uniform collection of air quality data various sample probe-siting criteria must be followed (see 40 CFR Part 58, Appendix E for details). These criteria are summarized below:

- The monitor probe shall be 2 to 15 meters above ground level.
- There must be unrestricted airflow 270° around the probe inlet or 180 ° if the probe inlet is on the side of a building.
- The probe must be at least 10 meters or further from the drip line of trees and optimally  $\geq 20$  meters.
- The sample line should be as short as practical and should be Perfluoroalkoxy (PFA) Teflon™ or equivalent. The sample line is replaced every two years.

## 7.3 Site Computer, DAS and Modem

The ECB is responsible for providing instrumentation and an operating system compatible computer, DAS, and a modem communication system. The operating system, data collection and storage capability, and remote communication system must be designed and acquired to maximize ease of data collection and transferring of data remotely. The entire data collection and communication system must be compatible with data retrieval programs operated by the database manager of the Ambient Monitoring Section. The ECB is encouraged to evaluate procedures, equipment, and systems to improve the data collection, data storage, and data transferring capabilities of the North Carolina NO<sub>2</sub> monitoring system.

## 7.4 Leak and Flow Checks

Leak and sample flow checks should be performed as part of the performance audit and during any standard maintenance of the T500U.

### 7.4.1. Detailed Pressure Leak Check

Obtain a leak checker that has a small pump, shut-off valve, and pressure gauge to create an over-pressure condition less than or equal to 10 psig. Alternatively, a tank of pressurized gas, with the two-stage regulator adjusted to  $\leq 10$  psig, a shutoff valve and a pressure gauge may be used. See section 11.3.5.1 of the manual for step-by-step instructions for completing the pressure leak check.

### 7.4.2 Sample Flow Check

Sample flow checks are useful for monitoring the actual flow of the instrument. A decreasing sample flow may point to slowly clogging pneumatic paths, most likely the sample flow restrictor or the sample filter. See section 11.3.5.2 of the manual for step-by-step instructions for performing a sample flow check.

## 7.5 Diagnostic Checks upon Installation

In order to ensure the monitoring equipment was not damaged in transit or during installation, ECB should initiate functionality testing for each sequence and each phase in a sequence. The installation functionality tests are intended to verify the instrument (and its associated components) have not suffered a catastrophic mishap from lab bench to field shelter. A functionality test is a test of the monitor, calibrator, cylinder, or zero air supply to evaluate whether the system is performing as expected. It may include running a zero and span or multiple points. This procedure is not a substitute for the initial calibration to be performed by the region.

## 7.6 Method Detection Limit (MDL) Determination

The MDL refers to the lowest concentration of a substance that can be reliably determined by a given procedure and typically is not provided by the vendor. A site-specific MDL establishes an estimate based on the routine

operation (and conditions) of the specific instrument in the network and provides a meaningful evaluation of data as it is aggregated across the precursor gas network. By establishing site specific MDLs, values less than the MDL can be flagged which would allow data users a more informed decision on the use of that data.

The MDL is determined by a series of low concentration standards and the average blank background. These aspects are input into the MDL procedure to establish the lowest concentration that is distinguishable from background with 99% confidence. ECB personnel will establish an MDL for the instrument initially and should be re-determined when changes to the instrument would reasonably expect to affect the sensitivity, such as replacing a detector lamp, detector and/or flow cell mirrors, for example. See DAQ-15-007.1 for detailed procedures for MDL determination for continuous gaseous criteria pollutant methods.

## 8.0 ROUTINE MAINTENANCE

---

### 8.1 Teledyne T500U NO<sub>2</sub> Analyzer

Periodic maintenance procedures should be performed when necessary to endure proper operation of the analyzer. Maintenance includes preventive, routine and corrective tasks. The ECB is expected to be entirely responsible for the corrective maintenance issues and to assist with preventative and routine maintenance that may fall outside the regions' comfort levels or capabilities. All maintenance activities must be documented by ECB personnel in the monitor's maintenance logbook. Step-by-step procedures for all maintenance activities should be followed as presented by the manufacturer in the instrument's operation manual (Chapter 11", "Maintenance" Version 0783D DCN6998 01 June 2015)

Items requiring maintenance by ECB include but are not limited to:

- Replacing the sample filter (annually or when the number of system hours is between 8750 and 9000 hours;
- Replacing the AREF Filter and Charcoal Filter (annually);
- Leak and Flow checks (performed after annual filter change and during performance evaluations); and
- Replacing the Internal Pump.

### 8.2 Teledyne Model T700U Calibrator

Periodic maintenance and/or adjustment of the calibrator is required to endure proper operation. In addition to mass flow controller and O<sub>3</sub> photometer re-certifications, which occurs every 12 months, the following maintenance activities are performed only when the calibrator malfunctions as determined by the site operator or ECB. Items requiring ECB maintenance are:

- Leak checking;
- Solenoid replacement and testing to determine where failures occurred;
- Circuit board replacement;
- Mass flow controller replacement;
- Internal adjustments; and
- Certification of mass flow controllers (reference SOP 2.3.5 for more details).

### 8.3 Model 701 Zero Air Generator Checks

Periodic maintenance and/or adjustment for the Zero Air Generator is required to ensure proper operation. The ECB shall re-certify the zero-air system once per year by:

- Verify the pressure gauge on the Zero Air Generator is reading  $30 \pm 2$  psig.



- Replacing chemicals that have been depleted (reference SOP 2.3.5 for more details).


## 9.0 ACCURACY AUDITS

Each analyzer must be audited by the ECB annually unless specified otherwise in the applicable QAPP (background monitoring sites operating CAPS monitors are audited quarterly). The audit must be performed using a calibrator and gas cylinder standard that is different from the standard and calibrator used for routine calibration and 1-pt QC checks. The audit calibrator must be certified against the primary flow standard every 12 months. The auditor must not be the operator who conducts the routine monitoring, calibrations, and analysis. No adjustments should be made prior to completing the audit. The monitor must operate in its normal sampling mode, and the concentration of the audit test gas and the concentration indicated by the analyzer is used to access the accuracy of the monitoring data. Several routine items that shall be included in the audit are:

- Security of the building;
- Site / building temperature (document NIST on the audit form); and
- Condition of the sample line, probe, and funnel (replace as needed) checked during each audit.

### 9.1 Audit Procedures

1. Allow the audit calibrator to equilibrate at least one-half hour before challenging the monitor
2. Using DAS software set the NO<sub>2</sub> channel flag to "Audit"
3. Connect the audit calibrator to the monitor and the zero-air system to the audit calibrator as per manufacturer's instructions (see Figure 3-15, page 45 of the instruction manual).



Rear Panel of the T700 Calibrator

4. Secure a separate certified Protocol cylinder of NO gas, connect and purge the regulator. Check for leaks.
5. Verify the audit calibrators' certification date recorded on the front panel label.



6. Using the AQ-121 form record site and audit device information as listed for each piece of equipment and gas cylinder.
7. Perform audit – at least 3 concentrations at Level 1, 5 and 6 and zero must be introduced to the analyzer by sequentially activating a GTPS and then GPT with the NO concentration set 10% greater than O<sub>3</sub> concentration. The generated O<sub>3</sub> ppb concentration is equivalent to the resulting actual concentration of NO<sub>2</sub>.
8. Complete an audit form (AQ-121) using the minute data collected during each audit point, recording the audit concentrations and five 1-minute averages.
9. The audit results should be within  $\pm 15\%$  of expected.
10. When complete, disconnect the audit calibrator, zero air system and audit cylinder. Reconnect site calibrator, zero air generator, and site cylinder as found upon arrival.
11. Set the NO<sub>2</sub> channel to “OK” and log out of site computer.

If the audit results are suspicious or unacceptable, the ECB supervisor or PPB staff will initiate an investigation of the problem and notify the responsible Regional Coordinator of the issue. Investigation of the audit results may include:

- Examination of the audit equipment;
- Review of the precision zero span records (both auto and manual); and
- Confirming the audit results with a follow-up audit.

The NO<sub>2</sub> performance audit is a “Pass-Fail” evaluation. The criteria for reevaluating the acceptance or failure of the NO<sub>2</sub> monitoring site are listed in the table below:

Nitrogen Dioxide Performance Audit Acceptance Limits		
Audit Level	NO <sub>2</sub> Audit Concentration	Site Monitor
Zero	0	< $\pm 1.5$ ppb
Level 1	2.5	< $\pm 1.5$ ppb or < $\pm 15.1\%$
Level 5	25	< $\pm 15.1\%$
Level 6	100	< $\pm 15.1\%$

## 10.0 REVISION HISTORY

This is a new SOP – there is no revision history.

## 11.0 REFERENCES

1. Model T500U CAPS NO<sub>2</sub> Analyzer Operation Manual (07834D DCN6998) 01 Jun 2015 (<http://www.teledyne-api.com/products/nitrogen-compound-instruments/t500u> )
2. NumaView™ Software Addendum to T-Series Analyzer Manuals (08296D DCN8039) (<http://www.teledyne-api.com/products/nitrogen-compound-instruments/t500u> )

## 12.0 APPENDICES

There are no appendices.