

Titration Standardization (NC WW/GW LC Policy 05/07/2019)

Titration standards prepared in the laboratory must be standardized initially and monthly; thereafter, unless a different frequency is specified in the referenced method. All certified titration standards which are purchased, may be used initially without standardization. The Certificate of Analysis must be kept on file. The certified titration standard must be standardized monthly; thereafter, for as long as it is used unless a different frequency is specified in the referenced method. If the normality (titer value) changes, a new titration standard at the specified normality must be used, or the sample results must be calculated using the newly determined normality. Quality Control standards do not take the place of titration standardization.